

Date July 03, 2012	Document C63.4-2009
-----------------------	------------------------

National Committee	Clause/Subclause	Paragraph Figure/ Table	Type of comment (General/ Technical/ Editorial)	COMMENTS	Proposed change	OBSERVATIONS OF THE SECRETARIAT on each comment submitted
	5	Figure 5.7	Technical	<p>We have a question on ANSI C63.4 2009, 5.7 figure for Clause 5.</p> <p>We would like to know how (to) interpret ANSI C63.4 2009, 5.7 figure for Clause 5. In that figure, page 23 of ANSI C63.4 2009, we interpret that the width of the ground plane needs to be at least 2m wider than the turn table. In an anechoic chamber, is that distance from the ground plane to the wall or from the ground plane to the tip of absorbent materials on the wall?</p>		<p>For alternative sites defined in clause 5.4.2 of C63.4:2009, figure 5 is suggestive (informative). There is no defined minimum size of ground plane and the distance between the EUT/Turntable and the absorber or walls is also undefined.</p> <p>The alternative site shall satisfy the volumetric site requirements (VNSA) defined in D.3.</p> <p>As stated in clause 5.4.3, Reflecting Ground Plane, figure 5 recommends a minimum configuration. Therefore there is no required minimum size of ground plane and the distance between the EUT/Turntable and the absorber or walls is also undefined. The alternative site has to satisfy the volumetric site requirements (VNSA) defined in D3. Considering the VNSA requirements are stringent, we recommend a distance of 1 m between the edge of the turntable and the front of the absorber. The final determination of compliance of an alternative test site is based on the volumetric site requirements.</p>